Operating Instructions

863 Valley View Road, Eighty Four PA 15330 USA
Tel: 724-941-9701 Fax: 724-941-1369 e-mail: skctech@skcinc.com
www.skcinc.com

Battery Pack for AirChek 52 Sample Pumps

To enhance battery life, SKC ships battery packs separate from the pump. Once installed, completely charge battery pack before operating the pump (see pump operating instructions). It may be necessary to charge the battery a few times before maximum battery capacity is achieved.

1. To remove a battery pack, loosen two security screws on bottom of pump case and allow them to fall to the out position.
2. Pinch battery pack release arms and pull the bottom half of the pump case away from the top half.
3. To install a battery pack, align arms on top half of case with slots on either side of the new battery pack housing. Ensure the back of the battery pack with the charging jack is on the same side as the belt clip.
4. Slide battery pack up until arms click into place.
5. Push each security screw into the battery pack housing and tighten.
6. Charge the new battery pack using the charging jack on the back of the battery pack.





Use of a repaired or rebuilt battery pack voids any warranty and the UL Listing for intrinsic safety.

Do not charge or operate pump from charger in hazardous locations!

Use only an SKC-approved charger and battery pack designed for the AirChek 52 Sample Pump to ensure reliable performance. Failure to do so voids intrinsic safety and any warranty.

Use of any device other than the approved battery pack to power the pump voids the UL Listing for intrinsic safety and any warranty.

Failure to follow warnings and cautions voids any warranty.

Notice: This operating instruction may not address all safety concerns (if any) associated with this product and its use. The user is responsible for determining and following the appropriate safety and health practices and regulatory limitations (if any) before using the product. The information contained in this document should not be construed as legal advice, opinion, or as a final authority on legal or regulatory procedures.

Form #40062 Rev 0907